

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)
Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 4.0 c
(Model: 4127, S/N: 210687448)

Traceable Thermometer 5.0 c
(Model: 4127, S/N: 192445446)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 5.0 c

(Model: 4127, S/N: 192445446)

Traceable Thermometer 4.0 c

(Model: 4127, S/N: 210687448)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 5.0 c

(Model: 4127, S/N: 210687885)

Traceable Thermometer 5.0 c

(Model: 4127, S/N: 192445446)

Validated by: Darren Lawrence, TS

Sign.:



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius) Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer - 64.61 c So Low Freezer Built-In Therm. - 64.7 c
(Model: 6411, S/N: 200336105) (Freezer Ext. Therm., M: PG214 03, S/N: 19386-4S14)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)
Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer -21.4 c
(Model: 4730, S/N: 181258814)

Traceable Thermometer -22.0 c
(Model: 6411, S/N: 200381437)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Weigh Scale Comparison Reading

Weight Scale, Henry Schein	176.2 LB	HealthOmeter	175.4 LB
-----------------------------------	-----------------	---------------------	-----------------

(S/N: HS-DBS03880)

(HDM519DQ7-15 N25452H)

Validated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Fahrenheit) Range within 1 Degree Fahrenheit

Oral Thermometer Comparison Reading

Welch Allyn Oral Thermometer
(Model: 690, S/N: 19351492)

98.8 F

Welch Allyn 99.1 F
(Model: 690, S/N: 19265085)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Fahrenheit)

Range within 1 Degree Fahrenheit

Oral Thermometer Comparison Reading

Welch Allyn Oral Thermometer 99.0 F

(Model: 690, S/N: 19450033)

Welch Allyn 98.8 F

(Model: 690, S/N: 19351492)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Fahrenheit) Range within 1 Degree Fahrenheit

Thermometer Comparison Reading

Welch Allyn Oral Thermometer 99.1 F

(Model: 690, S/N: 19265085)

Welch Allyn 98.8 F

(Model: 690, S/N: 19351492)

Validated by: **Darren Lawrence, TS**

Sign.:  _____

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Oral Thermometer Comparison Reading

Welch Allyn SureTemp +690 99.1 c

(Model: 690, S/N: 2029358)

Welch Allyn SureTemp + 690 98.8 c

(Model: 690, S/N: 19351492)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Ambient Thermometer 24.4 c

(Market Lab, S/N: 192684233)

Traceable Thermometer 24.0 c

(Model: 4127, S/N: 192445446)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

RPM Comparison Reading

Centrifuge RPM / 4500
(PowerSpin LX Model: C880, S/N: BT1903028)

Labcraft Digital Tachometer RPM /4502
(S/N: L100720)

Validated by: Darren Lawrence, TS
Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

B/P Comparison Reading

Omron BP 710N (139/99), 100 BPM
(S/N: 2011031458VG)

Omron BP 742N (136/99), 103 BPM
(S/N: 201811062282V)

Validated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

B/P Comparison Reading

Omron BP 742N (144/99), 102 BPM

(S/N: 2018u062281V)

Omron BP 710N (139/99), 100 BPM

(S/N: 2011031458VG)

Validated by: Darren Lawrence, TS

Sign.:



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

B/P Comparison Reading

Omron BP 742N (136/99), 103 BPM

(S/N: 201811062282V)

Omron BP 710N (139/99), 100 BPM

(S/N: 20111031458VG)

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Quincy Lab Incubator

(Model: 12-140, S/N: T-10134)

41.0 c

Vee Gee Brand Total Imm, (Mercury Glass Therm. No S/N visible)

39.0 c

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius) Range within 1 Degree Celsius

Thermometer Comparison Reading

So Low Freezer Built-In Therm. - 64.7 c
(Freezer Ext. Therm., M: PG214 03, S/N: 19386-4S14)
200336105)

Traceable Thermometer - 64.61 c
(Model: 6411, S/N:

Validated by: Darren Lawrence, TS

Sign.: _____



RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 3.91 c

(Model: 6411, S/N: 200381437)

Traceable Thermometer 4.0 c

(Model: 4127, S/N: 210687448)

Validated by: Darren Lawrence, TS

Sign.: _____

