

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Ambient Thermometer 24.4 c

(Market Lab, S/N: 192684233)

Traceable Thermometer 24.0 c

(Model: 4127, S/N: 192445446)

Validated by: Darren Lawrence, TS

Sign.: _____

