

Calibration Certificate

Certificate #: A1359941

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**

I.D.: 20267221

Manufacturer: WELCH ALLYN

Gage Type: THERMOMETER

Temp./RH: 22.8 C / 44.2 %

Cal Date: 8/23/2023

Department: CLINIC

Description: THERMOMETER - DIGITAL

Event Type: FIELD CALIBRATION

Serial Number: 20267221

Model Number: 901053

Performed By: 606

Location: FLOWER MOUND

Cal. Due Date: 8/23/2024

As Found Condition: In Tolerance

As Left Condition: In Tolerance

Calibration Notes

Tests points set by customer.

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	104.0	102.0	106.0	104.3		104.3	°F

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature

MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.