

Calibration Certificate

Certificate #: A1359949

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: LT 2009600
Manufacturer: HENRY SCHEIN
Gage Type: CENTRIFUGE
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: CENTRIFUGE
Event Type: FIELD CALIBRATIONSerial Number: LT 2009600
Model Number: C856
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tested at customer setting of 3500 RPM (Blood Range).

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	RPM	3346	3096	3596	3500		3500	rpm

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
MONARCH	ALD-347	TACHOMETER	7/10/2023	7/31/2024	A1341580

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCAA042	ROTATIONAL SPEED	1	7/7/2006

Technician Signature

MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.