

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

RPM Comparison Reading

Centrifuge RPM / 4500
(PowerSpin LX Model: C880, S/N: BT1903028)

Labcraft Digital Tachometer RPM /4502
(S/N: L100720)

Validated by: Darren Lawrence, TS

Sign.: _____

