

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakhri, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

rpm Comparison Reading

Centrifuge

(UNICO C808, S/N: F1702474)

Rpm / 3400

Labcraft Digital Tachometer / 3428

(S/N: L100720)

Validated by Darren Lawrence, TS

Sign.: 