

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Room Temperature Thermometer


Thermometer Comparison Reading (Fahrenheit)

Traceable Thermometer 22.5 c

(S/N: 6131)

Vee Gee Brand Total IMM 21.9 c

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Room Temperature Thermometer

Thermometer Comparison Reading (Fahrenheit)

Traceable Thermometer 70.7 F, Hum 26 %

(Model: 4040, S/N: 200597799)

Traceable Thermometer 70.7 F, Hum 26 %

(Model: 4040, S/N: 200770350)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)


**Freezer External Thermometer
Range within 1 Degree Celsius**

Thermometer Comparison Reading

Traceable Thermometer - 17.9 C Vee Gee Brand Total IMM - 18.2 C

(Model: 4730, S/N: 181258806)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

RPM Comparison Reading

Centrifuge, PowerSpin FX, C808 rpm/3400

(S/N: F1705234)

Tachometer rpm/3431

(Labcraft)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Pulse Oximeter Comparison Reading


Pulse Oximeter 96% O₂, 68 BPM

(M: MD300CN340, S/N: 20081858004877)

Accumed Pulse Ox 95% O₂, 69 BPM

(M: AC-CMS50D1, S/N: 21 0001979)

Validated by: Darren Lawrence, TS

Sign.:  _____

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Refrigerator External Thermometer


Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 3.0 C Vee Gee Brand Total IMM 3.2 C

(Model: 4127, S/N: 200668695)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Oral Thermometer Comparison Reading

WelchAllyn Sure Temp Plus, 98.2 F

(Model: 690, S/N: 2046899)

ReliOn, 98.3 F

(S/N: (01)00681131078214(10)0619)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Weigh Scale Comparison Reading


Detecto Weigh Scale 177.8 LB

(Model: PHR, S/N: C19013455)

Healthometer 178.0 LB

(Model: SN HDM 519DQ7-15 N25452H)

Calibrated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. F. Malik, Elgin, IL Site)

October 20, 2022

(Next Calibration Due: 10/20/2023)

Room Temperature Thermometer

Thermometer Comparison Reading (Fahrenheit)

Traceable Thermometer 70.7 F, Hum 26 %

(Model: 4040, S/N: 200770350)

Traceable Thermometer 70.7 F, Hum 26 %

(Model : 4040, S/N: 200597799)

Calibrated by: Darren Lawrence, TS

Sign.: 