

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

rpm Comparison Reading

Centrifuge

(UNICO C808, S/N: F1702474)

Rpm / 3400

Labcraft Digital Tachometer / 3428

(S/N: L100720)

Validated by Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Weight Comparison Reading

Health O Meter Professional 178.4 LB

(Model: 5000011352)

Health O Meter 178.4 LB

(5000068617)

Validated by: Darren Lawrence, TS

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakhri, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Weight Comparison Reading

Health O Meter Professional 178.4LB

(5000068617)

Health O Meter 178.4 LB

(Model: 5000013552)

Validated by: **Darren Lawrence, TS**

Sign.: 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 4.0 C Vee Gee Brand Tot. IMM 3.5 C

(4127, attached to frig, S/N: 200138948)

Validated by: Darren Lawrence, TS

Sign. 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer -24.0 C Vee Gee Brand Tot. IMM -23.1.0C

(4127, attached to freezer, S/N: 210687401)

Validated by: Darren Lawrence, TS

Sign. 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 25.7 C Vee Gee Brand Tot. IMM 25.7 C

(Room Thermometer, M: 4040, S/N: 210473162)

Validated by: Darren Lawrence, TS

Sign. 

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakih, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Fahrenheit)

Range within 1 Degree Fahrenheit

Thermometer Comparison Reading

Welch Allyn Oral Thermometer 98.1 F ReliOn 97.8 F

(Model: 690, S/N: 18244292)

(S/N: 0100681131 078214)

Validated by: Darren Lawrence, TS

Sign.: 