

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Quincy Lab Incubator

(Model: 12-140, S/N: T-10134)

41.0 c

Vee Gee Brand Total Imm, (Mercury Glass Therm. No S/N visible)

39.0 c

Validated by: Darren Lawrence, TS

Sign.: _____

