

RIA Consultant's, Inc.  
79235 Memphis Ridge  
Richmond, MI 48062

**Revive Research Institute Equipment Validation**

(Dr. F. Malik, Elgin, IL Site)

**October 20, 2022**

(Next Calibration Due: 10/20/2023)

**RPM Comparison Reading**

**Centrifuge, PowerSpin FX, C808 rpm/3400**

(S/N: F1705234)

**Tachometer rpm/3431**

(Labcraft)

**Calibrated by: Darren Lawrence, TS**

**Sign.:** 