

RIA Consultant's, Inc.  
79235 Memphis Ridge  
Richmond, MI 48062

## Revive Research Institute Equipment Validation

(Dr. Syed Raza, Lathrup Village, MI)

**June 5, 2023**

(Next Validation Due: 6/5/2024)

### Thermometer Verification (Celsius)

**Range within 1 Degree Celsius**

### **Thermometer Comparison Reading**

**SureTemp Plus 97.8 F**

(WelchAllyn 1690, S/N: 20297270)

**Reli-On 97.8 F**

(0619-07)

**Validated by: Darren Lawrence, TS**

**Sign:**

