

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Syed Raza, Lathrup Village, MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer -22.4 C Total IMM Thermometer -23.0 C

(Attached to freezer, S/N: 200668685)

Validated by: Darren Lawrence, TS

Sign.: 