

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Syed Raza, Lathrup Village, MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 24.7 C Total IMM Thermometer 24.6 C

(RM Thermometer, S/N: 210246132)

Validated by: Darren Lawrence, TS

Sign.:

