

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer 4.0 C Vee Gee Brand Tot. IMM 3.5 C

(4127, attached to frig, S/N: 200138948)

Validated by: Darren Lawrence, TS

Sign. 