

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius) Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer - 64.61 c So Low Freezer Built-In Therm. - 64.7 c
(Model: 6411, S/N: 200336105) (Freezer Ext. Therm., M: PG214 03, S/N: 19386-4S14)

Validated by: Darren Lawrence, TS

Sign.: _____

