

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revival Research Institute Equipment Validation

(Dr. Ali Moiin Dermatology, Troy, MI)

August 14, 2023

(Next Validation Due: 8/14/2024)

Thermometer Verification (Celsius)

Range within 1 Degree Celsius

Thermometer Comparison Reading

Traceable Thermometer -21.4 c
(Model: 4730, S/N: 181258814)

Traceable Thermometer -22.0 c
(Model: 6411, S/N: 200381437)

Validated by: Darren Lawrence, TS

Sign.: _____

