

Calibration Certificate

Certificate #: A1359929

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 200611954
Manufacturer: TRACEABLE
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 200611954
Model Number: 4127
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	10.7	8.7	12.7	10.0		10.0	°C

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359941



Calibration Performed By: 888-822-1299

ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235

For:

PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028

Equipment Information

I.D.: 20267221
 Manufacturer: WELCH ALLYN
 Gage Type: THERMOMETER
 Temp./RH: 22.8 C / 44.2 %
 Cal Date: 8/23/2023
 Department: CLINIC
 Description: THERMOMETER - DIGITAL
 Event Type: FIELD CALIBRATION

Serial Number: 20267221
 Model Number: 901053
 Performed By: 606
 Location: FLOWER MOUND
 Cal. Due Date: 8/23/2024
 As Found Condition: In Tolerance
 As Left Condition: In Tolerance

Calibration Notes

Tests points set by customer.

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	104.0	102.0	106.0	104.3		104.3	°F

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359939

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 20297389
Manufacturer: WELCH ALLYN
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 20297389
Model Number: 901053
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	104.0	102.0	106.0	103.9		103.9	°F

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359943

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 210472666
Manufacturer: TRACEABLE
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 210472666
Model Number: 4127
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	-21.9	-23.9	-19.9	-23.0		-23.0	°C

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359944

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 3P7044
Manufacturer: DETECTO
Gage Type: SCALES-5
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: SCALE
Event Type: FIELD CALIBRATIONSerial Number: NO SERIAL NUMBER
Model Number: 0-350LBS
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes****Test Points**

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Linearity 25% WC>=F	80.0	79.0	81.0	80.0		80.0	lbs
2	Linearity 50% WC>=F	175.0	174.0	176.0	175.0		175.0	lbs
3	Linearity 75% WC>=F	260.0	259.0	261.0	260.0		260.0	lbs
4	Linearity 100% WC>=F	350.0	349.0	351.0	350.0		350.0	lbs
5	Cornerload 1 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
6	Cornerload 2 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
7	Cornerload 3 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
8	Cornerload 4 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
9	Repeatability 1 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
10	Repeatability 2 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
11	Repeatability 3 175lbs	175.0	174.0	176.0	175.0		175.0	lbs

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
RICE LAKE WEIGHING	508B	WEIGHT	4/3/2023	4/30/2024	A1304250
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
RICE LAKE WEIGHING	606G	WEIGHT SET	2/14/2023	2/29/2024	A1282642

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCA503	SCALES	8	1/10/2011

Technician Signature

MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359948

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: K210201
Manufacturer: MAGEEK
Gage Type: BALANCES-5
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: BALANCE - PORTABLE
Event Type: FIELD CALIBRATIONSerial Number: K210201
Model Number: MM-TR-50-001-BK
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes****Test Points**

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Linearity 25%	12.500	12.499	12.501	12.500		12.500	gram
2	Linearity 50%	25.000	24.999	25.001	25.000		25.000	gram
3	Linearity 75%	37.500	37.499	37.501	37.500		37.500	gram
4	Linearity 100%	50.000	49.999	50.001	50.000		50.000	gram
5	Corner 1	25.000	24.998	25.002	25.000		25.000	gram
6	Corner 2	25.000	24.998	25.002	25.000		25.000	gram
7	Corner 3	25.000	24.998	25.002	25.000		25.000	gram
8	Corner 4	25.000	24.998	25.002	25.000		25.000	gram
9	Repeatability 1	25.000	24.999	25.001	25.000		25.000	gram
10	Repeatability 2	25.000	24.999	25.001	25.000		25.000	gram
11	Repeatability 3	25.000	24.999	25.001	25.000		25.000	gram

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
RICE LAKE WEIGHING	606A	WEIGHT SET - ANALYTICAL	12/20/2022	12/31/2023	A1262804
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCA502	BALANCES	8	5/15/2012

Technician Signature

MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359949

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: LT 2009600
Manufacturer: HENRY SCHEIN
Gage Type: CENTRIFUGE
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: CENTRIFUGE
Event Type: FIELD CALIBRATIONSerial Number: LT 2009600
Model Number: C856
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tested at customer setting of 3500 RPM (Blood Range).

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	RPM	3346	3096	3596	3500		3500	rpm

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
MONARCH	ALD-347	TACHOMETER	7/10/2023	7/31/2024	A1341580

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCAA042	ROTATIONAL SPEED	1	7/7/2006

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359929

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 200611954
Manufacturer: TRACEABLE
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 200611954
Model Number: 4127
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	10.7	8.7	12.7	10.0		10.0	°C

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359939

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 20297389
Manufacturer: WELCH ALLYN
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 20297389
Model Number: 901053
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	104.0	102.0	106.0	103.9		103.9	°F

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359941



Calibration Performed By: 888-822-1299

ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235

For:

PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028

Equipment Information

I.D.: 20267221
 Manufacturer: WELCH ALLYN
 Gage Type: THERMOMETER
 Temp./RH: 22.8 C / 44.2 %
 Cal Date: 8/23/2023
 Department: CLINIC
 Description: THERMOMETER - DIGITAL
 Event Type: FIELD CALIBRATION

Serial Number: 20267221
 Model Number: 901053
 Performed By: 606
 Location: FLOWER MOUND
 Cal. Due Date: 8/23/2024
 As Found Condition: In Tolerance
 As Left Condition: In Tolerance

Calibration Notes

Tests points set by customer.

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	104.0	102.0	106.0	104.3		104.3	°F

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359943

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: 210472666
Manufacturer: TRACEABLE
Gage Type: THERMOMETER
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: THERMOMETER - DIGITAL
Event Type: FIELD CALIBRATIONSerial Number: 210472666
Model Number: 4127
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tests points set by customer.


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Temperature	-21.9	-23.9	-19.9	-23.0		-23.0	°C

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
FLUKE	ASI-96	TEMPERATURE CALIBRATC	7/10/2023	7/31/2024	A1345079

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	CUSTOMER SPECS	CUSTOMER SPECIFICATIONS	8	6/23/2004
ALDINGER CO	PPCAA037-2	DIGITAL TEMPERATURE DEVICE	3	4/5/2016

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359944



Calibration Performed By: 888-822-1299

ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235

For:

PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028

Equipment Information

I.D.: 3P7044
 Manufacturer: DETECTO
 Gage Type: SCALES-5
 Temp./RH: 22.8 C / 44.2 %
 Cal Date: 8/23/2023
 Department: CLINIC
 Description: SCALE
 Event Type: FIELD CALIBRATION

Serial Number: NO SERIAL NUMBER
 Model Number: 0-350LBS
 Performed By: 606
 Location: FLOWER MOUND
 Cal. Due Date: 8/23/2024
 As Found Condition: In Tolerance
 As Left Condition: In Tolerance

Calibration Notes


Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Linearity 25% WC>=F	80.0	79.0	81.0	80.0		80.0	lbs
2	Linearity 50% WC>=F	175.0	174.0	176.0	175.0		175.0	lbs
3	Linearity 75% WC>=F	260.0	259.0	261.0	260.0		260.0	lbs
4	Linearity 100% WC>=F	350.0	349.0	351.0	350.0		350.0	lbs
5	Cornerload 1 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
6	Cornerload 2 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
7	Cornerload 3 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
8	Cornerload 4 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
9	Repeatability 1 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
10	Repeatability 2 175lbs	175.0	174.0	176.0	175.0		175.0	lbs
11	Repeatability 3 175lbs	175.0	174.0	176.0	175.0		175.0	lbs

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
RICE LAKE WEIGHING	508B	WEIGHT	4/3/2023	4/30/2024	A1304250
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
RICE LAKE WEIGHING	606G	WEIGHT SET	2/14/2023	2/29/2024	A1282642

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCA503	SCALES	8	1/10/2011

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359948

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: K210201
Manufacturer: MAGEEK
Gage Type: BALANCES-5
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: BALANCE - PORTABLE
Event Type: FIELD CALIBRATIONSerial Number: K210201
Model Number: MM-TR-50-001-BK
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes****Test Points**

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	Linearity 25%	12.500	12.499	12.501	12.500		12.500	gram
2	Linearity 50%	25.000	24.999	25.001	25.000		25.000	gram
3	Linearity 75%	37.500	37.499	37.501	37.500		37.500	gram
4	Linearity 100%	50.000	49.999	50.001	50.000		50.000	gram
5	Corner 1	25.000	24.998	25.002	25.000		25.000	gram
6	Corner 2	25.000	24.998	25.002	25.000		25.000	gram
7	Corner 3	25.000	24.998	25.002	25.000		25.000	gram
8	Corner 4	25.000	24.998	25.002	25.000		25.000	gram
9	Repeatability 1	25.000	24.999	25.001	25.000		25.000	gram
10	Repeatability 2	25.000	24.999	25.001	25.000		25.000	gram
11	Repeatability 3	25.000	24.999	25.001	25.000		25.000	gram

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
RICE LAKE WEIGHING	606A	WEIGHT SET - ANALYTICAL	12/20/2022	12/31/2023	A1262804
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCA502	BALANCES	8	5/15/2012

Technician Signature

MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.

Calibration Certificate

Certificate #: A1359949

**Calibration Performed By:** 888-822-1299ALDINGER CO
1440 PRUDENTIAL
DALLAS, TX 75235**For:**PRIME REVIVAL RESEARCH INSTITUTE, LLC
4101 KIRKPATRICK LANE
FLOWER MOUND, TX 75028**Equipment Information**I.D.: LT 2009600
Manufacturer: HENRY SCHEIN
Gage Type: CENTRIFUGE
Temp./RH: 22.8 C / 44.2 %
Cal Date: 8/23/2023
Department: CLINIC
Description: CENTRIFUGE
Event Type: FIELD CALIBRATIONSerial Number: LT 2009600
Model Number: C856
Performed By: 606
Location: FLOWER MOUND
Cal. Due Date: 8/23/2024
As Found Condition: In Tolerance
As Left Condition: In Tolerance**Calibration Notes**

Tested at customer setting of 3500 RPM (Blood Range).

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	*	As Left	Unit
1	RPM	3346	3096	3596	3500		3500	rpm

* In "As Left" column "A" denotes an adjusted test point, "F" a failed test point and "R" a report of value. Blank denotes an acceptable value.

Standards Used	I.D.	Description	Last Cal.	Cal. Due Date	Traceability #
TES	606BPTH	BAROMETER/HYGROMETE	7/27/2023	7/31/2024	A1349313
MONARCH	ALD-347	TACHOMETER	7/10/2023	7/31/2024	A1341580

Procedure Used	Procedure Name	Description	Revision Level	Revision Date
ALDINGER CO	PPCAA042	ROTATIONAL SPEED	1	7/7/2006

Technician Signature  MARCOS JIMENEZ

Aldinger Company certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). This report shall not be reproduced, except in full, without the written approval of Aldinger Company. The calibration interval has been specified by the customer. Any number of factors may cause the calibration to drift out of tolerance before the recommended interval has expired.