

RIA Consultant's, Inc.
79235 Memphis Ridge
Richmond, MI 48062

Revive Research Institute Equipment Validation

(Dr. Mona Fakh, Dbn. Hts., MI)

June 5, 2023

(Next Validation Due: 6/5/2024)

Thermometer Verification (Fahrenheit)

Range within 1 Degree Fahrenheit

Thermometer Comparison Reading

Welch Allyn Oral Thermometer 98.1 F ReliOn 97.8 F

(Model: 690, S/N: 18244292)

(S/N: 0100681131 078214)

Validated by: Darren Lawrence, TS

Sign.: 